TUCSON POLICE DEPARTMENT

CRIME LABORATORY BH

The undersigned Quality Assurance Specialist, having been qualified by the Arizona Department of Public Safety for Intoxilyzer 5000 inspection, and having performed such inspections on Intoxilyzer 5000 # 7, Serial # 64-001225, certifies by her signature:

on 1/26/04, with a result of 0.079 AC. 0.080 AC.

and on 2/19/04, with a result of 0.076 AC. 0.077 AC.

this instrument was tested with a known alcohol reference standard, and always produced results within the acceptable accuracy limits (from 0.070 AC. to 0.090 AC), as defined by R9-14-404.A.3.

No problems were reported concerning this instrument between these two dates.

Irene Doran

Irene Doran

ADPS QAS Permit #QAS 29586 (issue date 01/08/2004)

SUBSCRIBED AND SWORN to before me

this 20, day of February 2004

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC - ARIZONA
PIMA COUNTY
My Comm. Expires Sept. 4, 2004

Kathy A. Kinney

Notary Public

EXHIBIT Q

1 of 6
JD

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran / Misse
(Print Name)

INTOXILYZER SERIAL #(7)64 001225 LOCATION TRAFFIC
DATE Jan. 26 ~~19~~ 2004 TIME 1102
IED

JD Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

JD DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .362.

JD Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

JD Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

JD Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

JD Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

JD Alcohol free subject Test result 0.000 AC

JD Error recognition logic system functioning
Invalid test printed

JD Proper sample recognition system
Invalid sample printed
Deficient sample printed

N/A Completeness of sample purge with collection tube**

JD Calibration standard 0.080 AC
Results: 0.079 AC ; 0.080 AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Irene Doran Rhonda / Misse

0. JD alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
01/26/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:12
CAL. CHECK	.079	11:12
AIR BLANK	.000	11:13

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: TRAFFIC

OPERATOR: DORAN (46331)

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
01/26/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:14
CAL. CHECK	.080	11:14
AIR BLANK	.000	11:14

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: TRAFFIC

OPERATOR: DORAN (46331)

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI

Tucson Police Department Crime Laboratory

P. 386

p. 5 of 6

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 1/26/04 Time 1102
Intoxilyzer Serial# (7) 64-001225 Location TRAFFIC
QA Specialist Irene Doran / RMisse
(Print Name)

(JD) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(JD) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(JD) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(JD) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

(JD) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

(JD) 6. Push Start Test button or press enter on keyboard.

(JD) 7. Insert card in response to display.

(JD) 8. Air blank completed.

(JD) 9. Calibration check completed.
Test results 0.079 AC / 0.080 AC

(JD) 10. Air blank completed.

(JD) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(JD) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran Rhonda Misse

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

1076

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran (46331)
(Print Name)
INTOXILYZER SERIAL #(7)04-001225 LOCATION Traffic
DATE 2/19 ~~19~~ 2004 TIME 1100

Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .0363.

Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

Alcohol free subject Test result 0.000 AC

Error recognition logic system functioning
Invalid test printed

Proper sample recognition system
Invalid sample printed
Deficient sample printed

N/A Completeness of sample purge with collection tube**

Calibration standard 0.080 AC
Results: 0.076 AC & 0.077 AC

Instrument operating properly and accurately. YES✓ NO__

COMMENTS: _____

SIGNATURE Irene Doran

0. N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
02/19/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:08
CAL. CHECK	.076	11:08
AIR BLANK	.000	11:09

SUBJECT'S NAME

TRAFFIC
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
02/19/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:09
CAL. CHECK	.077	11:10
AIR BLANK	.000	11:10

SUBJECT'S NAME

TRAFFIC
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

386

586

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 2/19/04 Time 1100
Intoxilyzer Serial#(s) 04-001225 Location Traffic
QA Specialist Irene Doran (46331)
(Print Name)

1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
   Standard value: 0.080 AC
2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
4. Attach simulator to the simulator entrance port on the Intoxilyzer.
5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
6. Push Start Test button or press enter on keyboard.
7. Insert card in response to display.
8. Air blank completed.
9. Calibration check completed.
   Test results 0.076 AC ← 0.077AC
10. Air blank completed.
11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)