TUCSON POLICE DEPARTMENT

CRIME LABORATORY 184

The undersigned Quality Assurance Specialist, having been qualified by the Arizona Department of Public Safety for Intoxilyzer 5000 inspection, and having performed such inspections on Intoxilyzer 5000 # 7, Serial # 64-001225, certifies by her signature:

on 6/9/04, with a result of 0.078 AC, 0.076 AC,

and on 7/9/04, with a result of 0.074 AC, 0.075 AC,

this instrument was tested with a known alcohol reference standard, and always produced results within the acceptable accuracy limits (from 0.070 AC, to 0.090 AC), as defined by R9-14-404.A.3.

No problems were reported concerning this instrument between these two dates.

Jennifer L. Coombs

Jennifer L. Coombs

ADPS QAS Permit #QAS 29886 (issue date 06/15/2004)

SUBSCRIBED AND SWORN to before me

this 13, day of July, 2004.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC - ARIZONA
PIMA COUNTY
My Comm. Expires Sept. 4, 2004

Kathy A Kinney

Notary Public

186
ls

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000* ph

AGENCY Tucson Police QA SPECIALIST Irene Doran 46331
(Print Name)
INTOXILYZER SERIAL # 64001225 (#7) LOCATION Traffic
DATE June 9 ls 2004 TIME 1105

ls Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

ls DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "M". Reading is .335.

ls Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

ls Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

ls Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

ls Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

ls Alcohol free subject Test result 0.000 AC

ls Error recognition logic system functioning
Invalid test printed

ls Proper sample recognition system
Invalid sample printed ls
Deficient sample printed ls

N/A Completeness of sample purge with collection tube**

ls Calibration standard 0.080 AC
Results: 0.078AC + 0.076AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: replaced breath hose - breath hose not heating evenly

SIGNATURE Irene Doran

0.N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
06/09/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:05
CAL. CHECK	.078	11:05
AIR BLANK	.000	11:06

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: TRAFFIC

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC

CMI

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
06/09/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:06
CAL. CHECK	.076	11:07
AIR BLANK	.000	11:07

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: TRAFFIC

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC

CMI

Tucson Police Department Crime Laboratory

3 of 6

586

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 6/9/04 Time 1105
Intoxilyzer Serial# 64-001225(#7) Location Traffic
QA Specialist Irene Doran (46331)
(Print Name)

(ID) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(ID) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(ID) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(ID) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

(ID) (ID) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

(ID) (ID) 6. Push Start Test button or press enter on keyboard.

(ID) (ID) 7. Insert card in response to display.

(ID) (ID) 8. Air blank completed.

(ID) (ID) 9. Calibration check completed.
Test results 0.078+0.076 AC

(ID) (ID) 10. Air blank completed.

(ID) (ID) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(ID) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

P. 106
jlc

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES 154
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Jennifer Coombs 461160
(Print Name)

INTOXILYZER SERIAL # 64-001225 (#7) LOCATION ~~OTF~~ TRAFFIC
DATE July 9 ~~15~~ jlc 2004 TIME 1035

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .340.
✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".
✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".
✓* Clock time check. Mode selection switch S10 on or keyboard menu selection "E".
✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC
✓ Error recognition logic system functioning
Invalid test printed
✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓
N/A Completeness of sample purge with collection tube**
✓ Calibration standard 0.080 AC
Results: 0.074AC & 0.075AC

Instrument operating properly and accurately. YES ✓ NO __

COMMENTS: * moved clock ahead 1032 → 1036

SIGNATURE Jennifer L. Coombs

0. N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC #7
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
07/09/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	10:38
CAL. CHECK	.074	10:39
AIR BLANK	.000	10:39

SUBJECT'S NAME

TIME FIRST OBSERVED

TRAFFIC
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

#7

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
07/09/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	10:40
CAL. CHECK	.075	10:40
AIR BLANK	.000	10:40

SUBJECT'S NAME

TIME FIRST OBSERVED

TRAFFIC
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

P. 396

P. 5 of 6
JLC

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 7/9/04 Time 10.35
Intoxilyzer Serial# 64-001225 (#7) Location TRAFFIC
QA Specialist Jennifer Coombs 461160
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

(✓) 6. Push Start Test button or press enter on keyboard.

(✓) 7. Insert card in response to display.

(✓) 8. Air blank completed.

(✓) 9. Calibration check completed.
Test results 0.074AC & 0.075 AC

(✓) 10. Air blank completed.

(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Jennifer L. Coombs

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)